

HMK15 Humidity Calibrator



accepted and reliable method to calibrate humidity instruments. Usually two or three different salt solutions are used. Salts are chosen according to the application.

Certified salts

The HMK15 can be ordered with certified and pre-measured salts. A sample calibration is made from each batch in Vaisala's Measurement Standards Laboratory (MSL).

Calibrated thermometer

The HMK15 can be ordered with a thermometer, which is used for measuring the temperature during the calibration. It can also be used for checking the temperature measurement accuracy of the transmitter. The thermometer can contain either mercury (accuracy $\pm 0.3\text{ }^{\circ}\text{C}$ ($\pm 0.54\text{ }^{\circ}\text{F}$)) or red capillary fluid (accuracy $\pm 1.0\text{ }^{\circ}\text{C}$ ($\pm 1.8\text{ }^{\circ}\text{F}$)).

FINAS approved Measurement Standards Laboratory

Vaisala's Measurement Standards Laboratory is a FINAS accredited calibration laboratory. FINAS is a member of the EA (the European Cooperation for Accreditation).



In addition to laboratory use, Vaisala Humidity Calibrator HMK15 is suitable for on-site checks of humidity probes and transmitters.

Features/Benefits

- Easy and reliable calibration of humidity probes and transmitters
- Based on saturated salt solutions
- Fast temperature equilibration
- No external power required
- Suitable for laboratory use and on-site checks
- Chambers and transit covers make HMK15 easy to transport
- Pre-measured certified salts available
- Vaisala Service Centers offer accredited calibrations for humidity, temperature and barometric pressure.

No measuring instrument stays accurate by itself. It is essential that the functioning of an instrument is periodically checked against a reference. Vaisala has developed the Vaisala Humidity Calibrator HMK15 to make calibration and spot checking of humidity probes and transmitters easy and reliable.

Method used by leading laboratories

The operating principle of the HMK15 is based on the fact that a saturated salt solution generates a certain relative humidity in the air above it. The reading of the humidity probe or transmitter can then be adjusted accordingly. Many leading laboratories use this generally

Technical data

General

The standard HMK15 consists of the following parts:

- Two salt chambers, chamber covers and transit covers
- Base plate
- Choice of thermometers
 - mercury thermometer, Vaisala calibrated, order code 19728HM
 - thermometer with red capillary liquid, calibrated by manufacturer, order code 25130HM
- Measurement cup and mixing spoon

Options

Certified and ready dosed salts:		Order code:	Total uncertainty*:
LiCl salt	11 %RH	19729HM	(±1.3%RH)
MgCl ₂ salt	33 %RH	19730HM	(±1.2%RH)
NaCl salt	75 %RH	19731HM	(±1.5%RH)
K ₂ SO ₄ salt	97 %RH	19732HM	(±2.0%RH)
*Uncertainties given at +20 °C			
on exchanged water		19767HM	
Extra salt chambers		19766HM	
Carrying bag		HM27032	



For more information, visit www.vaisala.com or contact us at sales@vaisala.com

Ref. B210908EN-A ©Vaisala 2009
This material is subject to copyright protection, with all copyrights retained by Vaisala and its individual partners. All rights reserved. Any logos and/or product names are trademarks of Vaisala or its individual partners. The reproduction, transfer, distribution or storage of information contained in this brochure in any form without the prior written consent of Vaisala is strictly prohibited. All specifications — technical included — are subject to change without notice.

